



De Sio, C., Velthuis, J. J., Beck, L., Pritchard, J. L., & Hugtenburg, Rl. P. (2020). r-UNet: Leaf Position Reconstruction in Upstream Radiotherapy Verification. *IEEE Transactions on Radiation and Plasma Medical Sciences*.
<https://doi.org/10.1109/TRPMS.2020.2994648>

Publisher's PDF, also known as Version of record

License (if available):
CC BY

Link to published version (if available):
[10.1109/TRPMS.2020.2994648](https://doi.org/10.1109/TRPMS.2020.2994648)

[Link to publication record in Explore Bristol Research](#)
PDF-document

This is the final published version of the article (version of record). It first appeared online via IEEE at <https://ieeexplore.ieee.org/document/9094190>. Please refer to any applicable terms of use of the publisher.

University of Bristol - Explore Bristol Research

General rights

This document is made available in accordance with publisher policies. Please cite only the published version using the reference above. Full terms of use are available:
<http://www.bristol.ac.uk/red/research-policy/pure/user-guides/ebr-terms/>

r-UNet: Leaf Position Reconstruction in Upstream Radiotherapy Verification

C. De Sio, J. J. Velthuis, L. Beck, J. L. Pritchard, R. P. Hugtenburg

Abstract—Monolithic Active Pixel Sensor (MAPS) devices are an effective tool for upstream verification of Intensity Modulated Radiotherapy (IMRT) treatments. It is crucial to measure with high precision the positions of the multi leaf collimators (MLC) used to shape the beam in real time, in order to enhance the quality and safety of treatments. This work describes r-UNet, a deep learning based solution for leaf position reconstruction. The model is used to analyse the high-resolution images produced by a Lassena MAPS device in order to automatically determine the leaf positions. Image segmentation and leaf position estimation are performed simultaneously in a multi-task setting. r-UNet obtained an average Dice coefficient of 0.96 ± 0.03 for the reconstructed image masks in the held-out test set; whilst the mean squared error (MSE) resulting from the estimation of the MLC positions is 0.003 mm, with a resolution ranging between 45 and 53 μm for leaf extensions between 1 and 35 mm. On unseen leaf positions, r-UNet yielded a single-leaf resolution between 54 and 88 μm depending on the leaf extension, and an average MSE of 0.07 mm. These results were obtained using single frames of data collected at 34 frames per second.

Index Terms—Radiotherapy, Multi Leaf Collimator (MLC), Monolithic Active Pixel Sensors (MAPS), Deep Learning, Position Reconstruction, Image Segmentation.

I. INTRODUCTION

INTENSITY Modulated Radiotherapy (IMRT) is a technique used to treat cancerous tumours by shaping the photon beam through multi leaf collimators (MLC) [1], in order to deliver a concentrated radiation dose to the target region, while preserving the surrounding healthy tissue. As treatments get more complex, it is crucial to measure the delivered dose and the position of the collimator components (i.e. tungsten leaves) with high accuracy. MLC position calibration is usually performed on a monthly basis, and given the precision of standard calibration methods [2], a calibration tolerance of ± 1 mm for the MLC leaf edge position is currently reached [3]. From the evaluation of the effects of MLC positioning errors on the delivered treatments, it has been

suggested that the precision limit to maintain total dose errors below 2% is around 300 μm [4], [5]. As the calibrations are generally performed pre-treatment, a high-precision, real-time verification device would improve the treatment monitoring and could allow for faults to be detected during the treatment delivery and immediately addressed. This work describes an approach to MLC position reconstruction based on Artificial Intelligence (AI) techniques. In particular, we define r-UNet, a UNet-based model that is able to simultaneously detect the leaves in the images produced by MAPS sensors, and to estimate their position in a multi-task learning setting.

II. RADIOTHERAPY VERIFICATION

Several monitoring approaches exist to verify the radiotherapy treatment and measure the position of the MLC. The verification devices that are currently used can be divided into downstream and upstream detectors. The first are positioned behind the patient, and produce images that can be analysed to reconstruct the delivered dose and the shape of the treatment received by the patient. The most common devices are radiographic films and Electronic Portal Imaging Devices (EPIDs). The produced images, also known as portal images, are affected by beam distortion and scattering due to the presence of the patient.

Upstream detectors are placed before the beam reaches the patient. The produced image is therefore more clear compared to the previous approach, but the key challenge is keeping the beam attenuation to a minimum.

An alternative approach is using the log files produced by the LINAC itself as a measure of the MLC positions [6]. Use of log files is technically not verification, as it does not take into account the possibility that the MLC monitoring system is incorrectly calibrated or misaligned. Monolithic Active Pixel Sensor (MAPS) devices have been proven effective for upstream verification of radiotherapy treatments [7]–[9], as they can be made very thin and MAPS based systems can be produced with an attenuation below 1% [10].

A. Monolithic Active Pixel Sensors

A MAPS sensor consists of three layers (Fig.1). From bottom to top: a highly p-type doped substrate, referred to as the bulk; a lower p-type doped epitaxial layer (EPI) of silicon; and a highly p-doped layer on top of the EPI layer. In the top layer n-wells are created to collect the charge. The top layer also houses electronics, typically a single transistor where the first stage of amplification takes place. Sensors with more complicated circuits do exist.

C. De Sio is with the School of Physics, University of Bristol, H. H. Wills Physics Laboratory, Tyndall Avenue, BS8 1TL Bristol, United Kingdom. (e-mail: chiara.desio@bristol.ac.uk).

J. J. Velthuis is with the School of Physics, University of Bristol, BS8 1TL Bristol, U.K., also with School of Nuclear Science and Technology, University of South China, Hengyang 421001, China and with the Medical School, Swansea University, SA2 8PP, Swansea, U.K.

L. Beck and J. L. Pritchard are with the School of Physics, University of Bristol, H. H. Wills Physics Laboratory, Tyndall Avenue, BS8 1TL Bristol, U.K.

R. H. Hugtenburg is the Medical School, Swansea University, SA2 8PP, Swansea, U.K., also with the Department of Medical Physics and Clinical Engineering, Abertawe Bro Morgannwg University Hospital Board, Swansea SA2 4QA, U.K., and with the School of Physics, University of Bristol, BS8 1TL Bristol, U.K.




Fig. 1. Diagram of MAPS sensor. Incoming particles (photons) generate ionisation and electron-hole pairs are created. Electrons propagating in the silicon are then collected in the n-well.

When ionising radiation propagates through the sensor, electron-hole pairs are generated. The electrons produced in the epitaxial layer are confined to that layer by the built-in potential due to the different p doping concentrations. These electrons diffuse through the EPI layer, and eventually reach the depleted zone underneath the diode, in which they are collected.

The top layer that houses the transistors is very thin, much less than a micron thick. The top of the substrate layer is needed to create the built-in potential difference. The rest of the bulk is only for mechanical support. The epitaxial layer, where signal generation occurs, is typically between 2 and 20 μm thick. The device can thus be thinned from the back, up to 30 μm or less, without significant loss of signal-to-noise. In a radiotherapy verification scenario, a sensor with this thickness would attenuate the beam by less than 0.1%.

Previously, with the Achilles sensor [11] which has an area of $6 \times 6 \text{ cm}^2$ and a 15 μm pitch, a leaf edge resolution of $52 \pm 4 \mu\text{m}$ at the iso-centre has been achieved. This work was done using 0.1 s of data taken at 400 Monitor Units (M.U.) per minute, for leaves with a width of 1 cm at isocentre [10]. The algorithm used to calculate the leaf profile was based on a Sobel filter. Leaf misplacements as small as 0.5 mm were detected and moving leaves were tracked [12].

In this work we use the Lassena sensor [13]: a $12 \times 14 \text{ cm}^2$, 3T sensor with 50 μm pitch. Despite having larger pixel sizes, leading to a lower resolution, this device is 3-side butttable, which allows the coverage of large areas by tiling sensors in a $2 \times N$ configuration, without significant dead space in between. For example, a 2×2 matrix of these sensors can cover a large enough area to verify a radiotherapy treatment with field as large as $30 \times 30 \text{ cm}$, thus it could be clinically deployed in the future. Moreover, the sensor can be operated at a frame rate of 34 frames per second.

III. ARTIFICIAL INTELLIGENCE FOR RADIOTHERAPY VERIFICATION

Nowadays, machine learning (ML) is used for several applications in Medicine, Medical Imaging and Radiotherapy [14], [15]. In particular, different ML based solutions have been

defined in the literature for radiotherapy verification. For example, volumetric dose prediction [16] was performed using Fully Convolutional Neural Networks (FCNN) to predict new dose distributions from the analysis of existing treatment plans; portal images are used as input data to assess the quality of radiotherapy treatments, and classify the delivered treatment as acceptable or not acceptable, compared to the treatment plans [17]. ML models are also used to estimate the position of the MLC, using the positions reported in the LINAC log files as ground truth, as well as estimating the delivered dose [18].

In this work, images obtained using a Lassena sensor are used as input to a FCNN model that is trained to simultaneously detect the leaf in the sensor image, and extract its position (i.e. displacement from the leaf bank hosting the components of the MLC). From a ML perspective, this can be seen as the combination of an image segmentation problem and a regression problem. Once trained, the model produces a multiple output consisting of the segmented images and the estimated positions, inferred directly from corresponding input sensor images. The aim of this work is to achieve the best possible position resolution with the least possible amount of radiation, in order to be able to provide an early intervention. We are exploring ML techniques for their power in working with large datasets and providing a fast response (after the training process has been completed).

A. Data description and experimental setup

The Lassena device produces 2800×2400 pixel images, showing the position of the MLC leaves in terms of attenuation of the radiotherapy beam. Different regions are associated with different pixel intensities. A series of measurements was recorded keeping the sensor fixed in one position and moving a single central leaf of the MLC forward from the leaf bank in a $10 \times 10 \text{ cm}^2$ square field. The dataset used in this work is composed of approximately 9,000 images. For each considered leaf position (i.e. 1, 2, 3, 4, 10, 15, 20, 25, 30 and 35 mm) a sample of 900 consecutive frames was acquired with our sensor. The 900 frames correspond to ~ 30 seconds of data per position, acquired in 3 consecutive runs of 300 frames. During the acquisition, the MLC configuration and the detector were kept still. To avoid overfitting, as well as to test the generalisation capabilities of the model, two separate evaluation protocols have been adopted in this work. (A) to test the ability of the model to correctly estimate the position of a leaf from a single image, all the images corresponding to 4 (out of 10) positions, namely 1, 3, 15, 30 mm, have been used only for testing. In the rest of the paper, this dataset will be referred to as “external unseen dataset”. None of these images, as well as their corresponding leaf positions, have been used during the training of the neural network. The aim is to evaluate the ability of the model to generate correct estimations for “new” leaf positions never seen before, emphasising the advantages of the regression settings, over a more simplistic classification. (B) The data for the remaining 6 positions (i.e. 2, 4, 10, 20, 25, 35 mm, corresponding to 5,400 images) are further split into a training set and a “held-out” test set, accounting for the 80% and the 20% of this data set, respectively (i.e. 4,320

and 1,080 images). It is worth noting that the selection of the frames during the split is completely randomised (i.e. refers to randomly selected frames within the 30 seconds interval) so to avoid any possible bias in the selection of training and evaluation data. The held-out test set is kept completely unseen, whereas an additional 20% of the training set is further used for internal validation during the training process (i.e. 864 out of 4,320 samples). This internal validation set is used to analyse the evolution of the loss during the training, and to monitor possible cases of overfitting. The random selection algorithm used to sample images to account for the internal validation set is the same one used for the held-out test set generation. Stratification is used in the splitting, to keep the proportions of each leaf position in the three subsets and reduce biases in the training. A fixed seed is set in the random data splitting, for reproducibility purposes. The single leaf images used in this work have been collected at the Singleton Hospital in Swansea, using an Elekta Synergy LINAC to deliver a user-defined treatment at 400 M.U./min, in a 6 MV photon field. The MLC system is an Elekta Agility, featuring 160 tungsten leaves (2×80) with 5 mm width at iso-centre. The sensor was operated at 34 fps, collecting approximately 30 seconds of treatment per leaf position.

B. Ground truth masks generation

Binary masks are used as ground truth for the segmentation task. The mask generation steps are illustrated in Fig. 2. Input images (a) were initially processed to identify the contour of the leaf. The Chan-Vese segmentation algorithm [19] was used, as available in the scikit-image library [20]. From the obtained profile (b), the portion of the image containing the leaf was selected, and plotted on a black background, using the “fill_between” method from the matplotlib library [21]. This plot, showing a white leaf on a black background, is the binary mask (c). The mask is generated with the same size of the input image, so that it can be superimposed to filter the leaf as a segmentation mask. We considered single-leaf images due to the complexity in the generation of the ground truth masks.

IV. MODEL AND TRAINING

As the problem of identifying leaves in the sensor images requires the localisation of objects in the input data, a neural




Fig. 2. (a): input image; (b): contour selected by the Chen-Vese algorithm, superimposed to the input image; (c): ground truth binary mask used for training. The shown images refer to a leaf displacement of 25 mm.

network model inspired by UNet [22] was defined. It has been shown that UNet-based models are an effective technique for medical image segmentation, see for example [15], [23]. Our r-UNet model extends the classic UNet, including regression layers to estimate the MLC positions. This approach is end-to-end and multi-task, as the two estimations are produced simultaneously and directly from the single input images. The main advantage of this learning strategy is that the model is able to identify the leaves directly from raw data, and calculate their position without requiring any manual intervention (e.g. manually selecting the leaves). A schematic representation of r-UNet is shown in Fig. 3.

r-UNet is composed of a contracting path and an expansive path; its structure allows it to learn a segmentation of the input image, and to produce an output image of the same size (typically a binary mask), isolating the target structure. It is composed of convolutional blocks, each composed of two 2D Convolutions, each followed by a Rectified Linear Unit (ReLU) [24], with increasing or decreasing feature size depending on the path. The contraction is implemented by adding MaxPooling [25] downsampling layers after each convolutional block. Each convolutional block has double the number of feature channels of the previous one. Differently from the original UNet model, r-UNet uses smaller feature map sizes (i.e. 16, 32, 64, 128, 256). These sizes have been chosen experimentally, considering a trade-off between performance and computation time.

Each up-sampling step is performed with 2D transposed convolutions, and the number of feature channels is halved in the successive layers. Zero-Padding [25] is used to match the size of consecutive feature maps.

The two paths are also connected horizontally, concatenating every last layer of the convolutional blocks in the contracting path to the first layer of the corresponding block in the expansive path. This allows for the learnt features to be shared during the training of the model.

A fully-connected layer with sigmoid activation is added at the end of the contracting path to enable the MLC position estimation. The leaf position regression is performed considering the same feature embedding used in the expanding path to generate the output binary masks.

r-UNet is trained using the sensor images as input data, along with the binary masks containing the corresponding leaf region as ground truth for the segmentation, and the LINAC set leaf positions as target labels for the regression. Input images are converted to 8-bit, cropped, and rescaled by a factor 0.25 (size reduced to 350×350 pixels), to ease the processing and reduce the memory occupation. For each input image, r-UNet produces the pair (segmented image and estimated position). The performance for both tasks is evaluated separately and combined.

A. Loss functions

The r-UNet is trained on the 3,456 images of the training dataset, using a combined loss function. The Dice-Sørensen Coefficient [23], [26], [27] is used to optimise the leaf detection, whereas the leaf position regression is optimised by




Fig. 3. r-UNet, the multi-task FCNN model, composed of a contracting path (left-hand side), an expansive path (right-hand side), and additional regression layers (light green), to perform image segmentation and position estimation simultaneously. The dashed blue lines are used to picture the downsampling performed by pooling layers.

calculating the mean squared error (MSE) between target and predicted distances. The Dice Coefficient is a metric used to estimate the similarity (spatial overlap) between two sample sets (i.e. X and Y) [26]. It is defined as:

$$DSC = \frac{2|X \cap Y|}{|X| + |Y|} \quad (1)$$

The Dice Loss is expressed as $1 - DSC$ to be used as loss function.

The resulting combined loss is defined as:

$$L = \alpha \cdot Dice\ Loss + (1 - \alpha) \cdot MSE \quad (2)$$

The Adam optimiser has been used, with learning rate $lr = 10^{-4}$, $\beta_1 = 0.9$, $\beta_2 = 0.999$, $eps = 10^{-8}$ and no weight decay [28]. Different values for the α coefficient have been tested in the range $0.25 - 0.75$ with a 0.05 step, and performance have been evaluated for each trained model. The evaluation of the *best model* was conducted by considering the value of the combined loss on validation data, as well as the absolute difference between the average distance estimation on test data and on external unseen data. The best estimations were obtained for $\alpha = 0.4$, corresponding to $L = 0.4 \cdot Dice\ Loss + 0.6 \cdot MSE$. This particular choice allows both problems to be optimised, whereas other values led to model overfitting, e.g. low MSE in distance prediction but no (or poor) binary mask generated, or good quality mask but bad distance estimation. For example, for $\alpha = 0.75$ the leaf segmentation (Dice Loss) reached a plateau almost immediately (see Fig. 4). In this case no masks were generated.

On the other hand, although the average MSE on the held-out set seemed promising ($MSE = 0.005$ mm), the same model, tested on unseen positions led to an $MSE = 0.809$ mm. Thus the model was clearly overfitting the training




Fig. 4. Evolution of the Dice Loss for $\alpha = 0.75$. The loss function reaches a plateau around 0.0, indicating that the model is not learning.

positions. This further confirms the benefits of the multi-task method, in which the two tasks are bounded to be learnt simultaneously; i.e. the leaf position estimation is constrained by its corresponding leaf segmentation.

The best model was trained for 100 epochs, using an early stopping criterion. The evolution of the combined loss function across the training and validation phases is reported in Fig. 5, along with the evolutions of the Dice Loss and MSE. These two plots are shown separately due to difference in the corresponding scales. The peak observed in the Dice Loss evolution plot around epoch 55 may be due to the fact that the optimiser yielded an update to the learning rate to skip over potential saddle points. As can be seen from the figures, the training converges very quickly to a stable solution for the two learning tasks, with this coefficient value. To validate the conclusion, several starting points have been experimentally tested, leading to the same best model.




Fig. 5. Evolution of the loss functions during training (dashed lines) and validation (solid lines) of the best model ($\alpha = 0.4$). From top to bottom: combined loss, according to eq. 2; Dice Loss; MSE loss.

B. Model implementation

The model was implemented in Python using the PyTorch framework [29]. It was trained on a NVIDIA GeForce Titan Xp, producing a memory footprint of 2.5 GB, and performing 1 epoch in 4 minutes on the train dataset (composed by 3456 images in training and 864 in validation). The inference time was 0.17 s per image (1,080 in the test dataset and 3,532 images in the external unseen dataset).

V. RESULTS

A. Performance on test dataset

The trained r-UNet model was tested on the 1,080 images, corresponding to 20% of the total dataset, not used for training. The Dice Coefficient resulting on test set images is 0.96 ± 0.03 , whereas the average MSE on the same images is 0.003 mm. The distribution of the prediction error (evaluated as difference between estimated and true leaf extension) for the test dataset is shown in Fig. 6. Calculated as the σ of the Gaussian fit, the position resolution at iso-centre for all test data is $50 \pm 1 \mu\text{m}$, ranging from $45 \mu\text{m}$ to $53 \mu\text{m}$ depending on the leaf extension, as shown in Fig. 9.

An example of the binary mask estimation from an input image as produced by the r-UNet is shown in Fig. 7.

B. Performance on external unseen dataset

To assess the generalisation capability of the proposed approach, the model performance was tested on four samples




Fig. 6. Error distribution on all test dataset and gaussian fit.




Fig. 7. (a): input test image; (b): r-UNet-generated binary mask.

of unseen positions: 1, 3, 15, 30 mm. Notably, r-UNet was not shown these positions in training. The obtained results are good: the average MSE ranges from 0.01 to 0.10 mm, depending on the positions. The distribution of the prediction errors for a sample of unseen images obtained for a leaf extension of 15 mm is shown in Fig. 8. A Gaussian fit yielded a reconstruction resolution of $54 \mu\text{m}$ and a MSE of 0.07 mm.




Fig. 8. Distribution of reconstructed position error for on a sample of unseen data (15 mm) and gaussian fit, yielding a reconstruction resolution of $54 \mu\text{m}$ and a MSE of 0.07 mm.

The single-leaf resolutions, calculated for each estimated position (held-out test and external unseen datasets) are shown in Fig. 9. Overall, the position resolutions range between 54 and $88 \mu\text{m}$ and are slightly worse in the case of unseen

positions, but still below $100\ \mu\text{m}$ even in the worst case (3 mm leaf extension). Furthermore, the leaf position resolution does not depend on the leaf extension.

These leaf position resolutions are far below the $300\ \mu\text{m}$ precision required to limit dose errors to below 2%. Please note that these results are achieved in single frames which are taken at 34 frames per second. The speed and precision allows for real time intervention during treatments.

In figure 10, the estimated positions are plotted as a function of the LINAC set positions. A straight line fit resulted in a slope of 0.999 ± 0.004 and an intercept of 0.061 ± 0.032 , which is compatible with the expected relation of $y=x$. Hence, the procedure yields the correct positions with excellent precision.




Fig. 9. Position resolution at iso-centre for the test data (blue) and unseen data (black).




Fig. 10. Estimated vs set leaf positions, with linear fit compatible with $y = x$.

C. Uncertainty in the reference positions

The “intended” leaf positions, i.e. the MLC positions set in the Linac, were used as ground truth to train the network and evaluate the performance of the estimations. The data was collected in small consecutive sessions, to reduce the possibility of leaf repositioning by the machine. For each configuration (i.e. single leaf displacement), 3 runs of 300 frames were taken, corresponding to ~ 10 seconds each, without changing

the leaf position or moving the sensor between consecutive runs. The stability of the detected leaf positions was analysed as a function of the frame number, to assess whether a (major) leaf repositioning occurred during the data taking.




Fig. 11. Estimated position as a function of the frame number for the 15mm case. The 3 highlighted regions indicate the central 200-frame subsets selected to study the stability.




Fig. 12. Residual distributions for 15mm, shown for the 3 regions separately. The corresponding resolutions are $51\ \mu\text{m}$, $53\ \mu\text{m}$ and $54\ \mu\text{m}$, respectively. The dashed black line, indicating the mean of the central distribution (blue, region 2), is shown for reference.

The results are shown in Fig. 11 for the 15 mm case. The estimated positions appear to be flat throughout the whole




Fig. 13. Average of estimated position for single-leaf configuration, shown for each region, separately. The estimations in each consecutive data acquisition run are compatible with one another. Corrected (shifted) positions are shown to ease the visualisation, with the corresponding leaf position in mm above.

~30 s data taking. This was confirmed by selecting a subset of 200 frames for each acquisition and calculating the residual distributions for each region, see Fig. 12. Finally, Fig. 13 shows that no significant repositioning took place during our data taking: for each leaf position, the average estimated positions in the 3 regions are compatible with one another. Hence, the obtained resolution is a good measure of the uncertainty on the leaf position.

VI. TOWARDS MULTI-LEAF RECONSTRUCTION FOR RADIOTHERAPY VERIFICATION

The results shown in this work were obtained for a 6 MV flattened radiotherapy beam (FF) for a single central MLC leaf displaced forward in a 10×10 cm² photon square field. Single leaf resolutions were calculated for 10 different displacements and the obtained results demonstrate the power of the novel approach, which is based on ML. However, the signal in the sensor, and thus the reconstructed leaf position, does not only depend on the position of the leaf under study but also on the position of its neighbours and the field configuration. Now that it has been proven that the approach presented here works very well for the simple configurations used here, we plan to extend this approach to more complicated scenarios and study the dependency of the leaf resolution on different field sizes, different beam energies, as well as using Flattening Filter Free (FFF) beams. We will progress to the study of multiple leaf detection and complex radiotherapy treatments with moving leaves. In essence, leaf position reconstruction is about detecting the transition between the covered part of the sensor, which will be in the dark, and the uncovered part, where the signal will be high(er) due to the beam. The electron field smears this transition to a certain extent. Changing the field size results in a different contribution due to the electron field but as the rapid light-dark transition due to the partially blocked photon beam is still present, the algorithm will still detect the leaf position. The different electron field might lead to a different slope of this transition and thus to a different leaf position resolution, but the technique will still work and the effect should be minimal. Similarly, a flattening filter free (FFF) beam will mainly change the signal in the region where the sensor is not covered by the leaves as its amplitude will become dependent on the leaf position, but the rapid transition will still be present. Hence, we do not expect a major difference in leaf position resolution for different fields, beam energies and FFF beams. As for the detection of multiple leaf

positions, we will use a region of interest approach. Once the regions of interest have been identified (image segmentation), the estimation will be performed for each identified leaf iteratively, taking into account the estimated positions for the neighbouring leaves. Applying the technique to treatments with moving leaves like VMAT is more challenging, partly because it is difficult to define a reference position to compare to for a moving leaf. However, our sensor integrates images at 34 frames per second. Hence, the maximum movement between consecutive frames is about 1.47 mm for a Varian linac [30] and 1.91 mm for a Elekta Agility [31]. Since the detector integrates between readout cycles, the light-dark transition will be smeared out which will give information on the leaf position and average leaf speed. This information can be used in training to obtain better estimates. This will be the subject of further study.

VII. CONCLUSION

Now that IMRT is widely in use and with the novel trends towards more intense treatments, it becomes more and more important to independently and dynamically verify the positions of the MLC. Here a machine learning algorithm is presented which reconstructs the leaf positions in real time with a precision well below 100 μ m. All the estimations are obtained on a single frame basis, which corresponds to 0.03 s of treatment. The resolutions are well below the proposed limit of 300 μ m, which in turn results in dosimetric errors below 2%. Multiple leaf position reconstruction is currently under development.

ACKNOWLEDGMENT

This work was supported by the Science and Technology Facilities Council [ST/T002646/1].

REFERENCES

- [1] M. Jeraj, V. Robar, *Multileaf collimator in radiotherapy*, Radiol Oncol 2004; 38(3): 235-40.
- [2] M. N. Graves, A. V. Thompson, M. K. Martel, D. L. McShan, B. A. Fraass, *Calibration and quality assurance for rounded leaf-end MLC systems*, Medical Physics 28(11):2227-2233, 2001.
- [3] E. E. Klein, J. Hanley, J. Bayouth, F. F. Yin, W. Simon, S. Dresser, and C. Liu (2009). Task Group 142 report: Quality assurance of medical accelerators a. Medical physics, 36(9Part1), 4197-4212.
- [4] A. Rangel and P. Dunscombe, *Tolerances on MLC leaf position accuracy for IMRT delivery with a dynamic MLC*. Medical physics, 36(7), 3304-3309, 2009.

- [5] A. Agarwal, N. Rastogi, K. M. Das, S. A. Yoganathan, D. Udayakumar, R. Naresh and S. Kumar, *Evaluating the dosimetric consequences of MLC leaf positioning errors in dynamic IMRT treatments*, Journal of Radiotherapy in Practice, 1-7, 2019.
- [6] Monitoring daily MLC positional errors using trajectory log files and EPID measurements for IMRT and VMAT deliveries, A. Agnew et al., Physics in Medicine & Biology, Volume 59, Number 9, 2014.
- [7] J. J. Velthuis, R. P. Hugtenburg, D. Cussans, M. Perry, C. Hall, P. Stevens, and A. McKenzie. *The VANILLA sensor as a beam monitoring device for X-ray radiation therapy*, Applied Radiation and Isotopes, 83, 8-11, 2014.
- [8] J. J. Velthuis, R. P. Hugtenburg, C. Hall, R. Page, & P. Stevens, *Monitoring of a therapeutic X-ray beam using an active pixel sensor detector*. Patent granted in EU on 04/04/2018, EP2654895B1.
- [9] J. J. Velthuis, R. P. Hugtenburg, C. Hall, R. Page, & P. Stevens, (2016). *Upstream direct x-ray detection*. U.S. Patent No. 9,517,358. Washington, DC: U.S. Patent and Trademark Office.
- [10] R. F. Page et al., *Using a Monolithic Active Pixel Sensor for Monitoring Multileaf Collimator Positions in Intensity Modulated Radiotherapy*, IEEE Transactions on Nuclear Science vol 61, 2014.
- [11] N. Guerrini, R. Turchetta, G. Van Hoften, R. Henderson, G. McMullanc and A. R. Faruqic, *A high frame rate, 16 million pixels, radiation hard CMOS sensor*, Journal of Instrumentation 6, 2010.
- [12] L. Beck, J. J. Velthuis, S. Fletcher, J. A. Haynes and R. Page, *Using a TRAPS upstream transmission detector to verify multileaf collimator positions during dynamic radiotherapy delivery*, Applied Radiation and Isotopes, 108951, 2019.
- [13] I. Sedgwick et al., *LASSENA: A 6.7 Megapixel, 3-sides Buttable Wafer-Scale CMOS Sensor using a novel grid-addressing architecture*.
- [14] V. Weidlich, G. A. Weidlich, *Artificial Intelligence in Medicine and Radiation Oncology*. Cureus, 10(4), e2475, 2018.
- [15] G. Litjens et al., *A survey on deep learning in medical image analysis*, Medical Image Analysis, 42, 60 - 88, 2017.
- [16] V. Kearney, J. W. Cham, G. Valdes, T. Solberg, S. S. Yom, *The application of artificial intelligence in the IMRT planning process for head and neck cancer*, Oral Oncology, Vol. 87, pages 111-116, December 2018.
- [17] K. Leszczynski et al, *Application of a fuzzy pattern classifier to decision making in portal verification of radiotherapy*, Phys. Med. Biol. 44 253, 1999.
- [18] J. N. Carlson, et al., *A machine learning approach to the accurate prediction of multi-leaf collimator positional errors*. 2016 Phys. Med. Biol. 61 2514, 2016.
- [19] R. Cohen, *The Chan-Vese Algorithm*, arXiv e-prints, Jul. 2011, arXiv:1107.2782.
- [20] S. Van der Walt et al., *scikit-image: image processing in Python*, PeerJ 2-6, 2014.
- [21] J. D. Hunter, *Matplotlib: A 2D graphics environment* Computing in Science & Engineering, vol. 9, no. 3, 90-95, 2007.
- [22] O. Ronneberger et al., *U-Net: Convolutional Networks for Biomedical Image Segmentation*, MICCAI, 2015.
- [23] F. Milletari et al., *V-Net: Fully convolutional neural networks for volumetric medical image segmentation*, IEEE International Conference on 3D Vision, 2016.
- [24] V. Nair, G. E. Hinton, *Rectified Linear Units Improve Restricted Boltzmann Machines*, ICML'10, 807-814, 2010.
- [25] V. Dumoulin, F. Visin, *A guide to convolution arithmetic for deep learning*, 2016
- [26] K. H. Zou, et al. *Statistical validation of image segmentation quality based on a spatial overlap index*, Academic radiology vol. 11,2 (2004): 178-89. doi:10.1016/s1076-6332(03)00671-8.
- [27] C. H. Sudre et al., *Generalised Dice overlap as a deep learning loss function for highly unbalanced segmentations.*, Deep Learning in Medical Image Analysis and Multimodal Learning for Clinical Decision Support LNCS, 2017.
- [28] D. P. Kingma and J. Ba, *Adam: A Method for Stochastic Optimization*, arXiv, 2014.
- [29] A. Paszke, S. Gross, S. Chintala, G. Chanan, E. Yang, Z. DeVito, Z. Lin, A. Desmaison, L. Antiga, A. Lerer, *Automatic differentiation in PyTorch*, 2017.
- [30] T. Y. Lim, I. Dragojević, D. Hoffman, E. Flores-Martinez, and G. Y. Kim, *Characterization of the HalcyonTM multileaf collimator system.*, J Appl Clin Med Phys, 20: 106-114. doi:10.1002/acm2.12568, 2019.
- [31] V. P. Cosgrove, C. Thompson, D. Chrisophides et al. (2009) *Physical characterization of a new concept design of an Elekta Radiation Head with integrated 160-leaf multi-leaf collimator*. Poster presented at ASTRO 2009.